

DECLASS REVIEW by NIMA/DOD

7 September 1961

MEMORANDUM FOR: Chief, TP&DS

SUBJECT:

Trip Report (Los Angeles, California, 21-31

August 1961)

25X1A

- 1. The undersigned, along with of this office 25X1A visited the for the purpose of monitoring three current contracts: (1) the HTA-3 Wide Film Processor; (2) the Image Quality Meter; and (3) the Image Enhancement Device. The first three days of this period were spent with at whica5X1A time the Image Quality Meter and Image Enhancement Device were extensively discussed. The 13th and 14th Monthly Progress Report, Image Quality Meter, Contract No. and the25X1A 14th and 15th Monthly Progress Report. Image Enhancement Program, contract No. are enclosed.
- 2. The Image Quality Meter is nearing completion and some preliminary tests were conducted on this instrument. The values obtained for acuteness and granularity are fairly valid determinations provided the proper imagery is selected. The values obtained for the resolution reading will require considerable operator judgement as to the type of imagery and the density and the physical location within the format area. This instrument is scheduled for delivery early this fall.

25X1A

25X1A